

Dear Mr Hadzi-Djordjevic,

I want to confirm that I had read Your letter „Questions“ and it is my obligation as the project manager to send you next answers:

1. Regarding your first question, i.e. the statement that „The tender refers to the Rapid Resolution Chromatography while specifications are made for Ultra Performance High Pressure Liquid Chromatograph UHPLC“, I can add that under Rapid Resolution Chromatography we meant the Liquid Chromatograph (LC) enabling rapid and high-quality resolution of complex mixtures with characteristics quoted in the tender material. These characteristics are fulfilled by at least two instruments of different manufacturers. However, in order to leave a possibility for participation of other manufacturers whose LC system can afford minim. 1200 bar, we decided to correct the required characteristics of the pump in this way:

**Binary or quaternary pump, pressure min. 1200 bar, max. flow 5 mL/min.**

2. In answering to your second question, I may say that we did not precise that all detectors must be from the same manufacturer. It is understood that the software following the LC must be compatible with all detectors and control the functioning of the instrument as a whole. Also, compatibility of all modules must be guaranteed – otherwise, the problems which might arise as the result of a possible incompatibility (for example, communication problems between the modules of different manufacturers) have to be solved on the expense of a deliverer. In addition, a candidate must be able to satisfy all other specified requirements regarding maintenance of the full equipment.

U Novom Sadu,  
July 28, 2011



Prof. Dr. Biljana Abramović  
Project manager