

Belgrade 22nd July 2011

University of Novi Sad,
Faculty of Sciences:
Prof. dr Biljana Abramović
Trg Dositeja Obradovića no. 3
21000 Novi Sad
E-mail: biljana.abramovic@dh.uns.ac.rs

Dear Prof Abramovic,

PUBLICATION REF.: HU-SRB/0901/121/116/005 – supply LC

The reference is made to the above mentioned tender. Please be so kind and answer on the following questions:

1. The tender refers to the Rapid Resolution Chromatography while specifications are made for Ultra performance High Pressure Liquid Chromatograph UHPLC. The specification of the UHPLC pump is based on one manufacturer particularly Agilent model 1290. Our Principal waters Corporation invented UHPLC, but due to the strict technical specification, we are not in position to participate. Requested specification for pressure of min. 1200 bar is not critical for performing chromatography in UHPLC or RR mode. Technical specifications must afford equal access for candidates and renderers and should not have the effect of creating unjustified obstacles which will eliminate certain products or economic operators by directly favoring a certain manufacturer and its instruments.


Question:

In order to make this procurement procedure with equal access and non favorable kindly confirm that Binary or Quaternary UHPLC pump with min. pressure up to 1000 bar is acceptable.

2. Question: Can we offer detectors from different manufacturers or all components must be from the single manufacturer same as UHPLC pump?

Your positive reply will be very much appreciated.

with kindest regards,



Ljubomir Hadzi-Djordjevic